The Association of Test Houses for Fume Cupboard Testing in ENLa in DIN (VPLD)

CERTIFICATE

It is hereby confirmed that the

Fume Cupboard T120

of the manufacturer

Renggli AG
Industrie-Ost
Birkenstrasse 31
6343 Rotkreuz
Switzerland

was tested according to

EN 14175
part 3

at the total volume flow rate of **480 m³/h**,
with a nominal face velocity of **0.34 m/s**
With this exhaust volume flow rate the requirements of
- **BG Chemie, Germany, issued 29.07.2003**
are completely fulfilled

The details and data of the complete test procedure
are content of the type test report

FCS/EN/93/2006
which is given to the manufacturer on
12. June 2006

Tested by:
Tintschl BioEnergie und
Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen

.....
Dipl.-Ing. Bernd Schubert
Erlangen, 11.07.2014

Tintschl ®

CERTYFIKAT

Niniejszym potwierdza się, że

Wyciąg T120

Producenta:

Renggli AG
Industrie-Ost.
Birkenstrasse 31
6343 Rotkreuz
Szwajcaria

Został przetestowany zgodnie z

o łącznej wartości natężenia przepływu 480 m³ /h, o nominalnej prędkości powietrza 0,34 m/s

Z niniejszą szybkością przepływu objętości wydechowej, wymagania - BG Chemie, Niemcy, wydane 29.07.2003 zostały całkowicie spełnione

Szczegółowe informacje i dane o pełnej procedurze badania są zawartością typu sprawozdania z testu

FCS/EN/93/2006

których oddano dla producenta w dniu 12 czerwca 2006